


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	PICO70 Arterial Blood Sampler
Lot No.	956-547
Manufacturer	Radiometer
Country of Origin	Denmark
Reference	https://healthycanadians.gc.ca/recall-alert-rappel-avis/hc-sc/2021/75381r-eng.php
Device picture	
Reason of Recall	NHRA initiates this FSN due to an issue with PICO70. The issue relates to the sterile barrier system for the product. During internal stress testing on PICO70 Arterial Blood Samplers a breach in the sterile barrier system was observed in one sample.
Action should be taken	Please stop using the above mentioned defected medical device and contact the authorized representative Wael Pharmacy at vincent@waelpharmacy.com to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh