

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	BD Veritor™ Plus Analyzer
Effected Device	256066
Lot No.	Unexpired Analyzers up to 31/08/2024
Manufacturer	BD
Country of Origin	USA
Reference	https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfres/res.cfm?id=191328
Reason of Recall	NHRA initiates this FSN due to a fire resultant from power issues could cause the user 2nd or 3rd degree burns and/or smoke inhalation should the user have been close enough to the device. Additionally, fire could result in property damage.
Action should be taken	Please stop using the above mentioned medical device and contact the authorized representative Wael Pharmacy Co. W.L. L at vincent@waelpharmacy.com to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh