


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	CITOTEST Transport Medium
Serial No.	2118-0019
Lot No.	042001
Manufacturer	Citotest Labware Manufacturing Co.,Ltd
Country of Origin	China
Reference	https://ncmdr.sfda.gov.sa/Secure/CA/CAViewRecall.aspx?caid=4&rid=15284
Device picture	
Reason of Recall	NHRA initiates this FSN because the virus specimens transportation tube may leak.
Action should be taken	Stop using the above-mentioned defected device and contact the supplier to take the necessary action for replacement.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh