



Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	CarbonAid Co2 Diffuser
Lot No.	205330
Manufacturer	Cardia Innovation AB
Country of Origin	Sweden
Reference	https://ncmdr.sfda.gov.sa/Secure/CA/CaViewRecall.aspx?caid=15&rid=16098
Picture	<div style="display: flex; justify-content: space-around; align-items: center;"> <div style="text-align: center;">  <p>Original Label</p> </div> <div style="text-align: center;">  <p>Counterfeited</p> </div> </div>

For more information please contact Medical_Devices@nhra.bh

Reason of Recall	<p>NHRA initiates this FSN due to SFDA has issued an alert stating the existence of counterfeited devices in the market.</p> <p>The label of the counterfeited device:</p> <ul style="list-style-type: none">• Doesn't contain product number.• Symbols of storage and temperature are not available• Lot number not recognized by legal manufacture
Action should be taken	<p>Please stop using the above mentioned medical device and contact the authorized representative or the manufacturer to take the necessary action for recall.</p>

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh