


## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details		
<b>Device Name</b>	Contact Lenses Solutions	
<b>Device Model</b>	<b>Model Name</b>	<b>Lot No.</b>
	RENU Multiplus Multi-purpose 360ml	ME2970, MF5135, MF1656, ME3081
	RENU Multiplus Multi-purpose 120ml	ME3124, ME4078
	Starter Kit RENU Multiplus Multi-purpose 60ml	ME3124, ME3848, MF1849
	BIOTRUE Multiplus Multi-purpose 300ml	ME3693
	BIOTRUE Multiplus Multi-purpose 120ml	ME2438
	Starter Kit BIOTRUE Multiplus Multi-purpose 60ml	MF1642
<b>Manufacturer</b>	Bausch and Lomb INC	
<b>Country of Origin</b>	Italy	
<b>Reference</b>	<a href="https://bausch.co.uk/news/bausch-lomb-announces-voluntary-recall#:~:text=LAVAL%2C%20Quebec%2C%20May%2027%2C,conditioning%20solution%2C%20Boston%2CAE%20Simplus">https://bausch.co.uk/news/bausch-lomb-announces-voluntary-recall#:~:text=LAVAL%2C%20Quebec%2C%20May%2027%2C,conditioning%20solution%2C%20Boston%2CAE%20Simplus</a>	

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)

<p><b>Device picture</b></p>	
<p><b>Reason of Recall</b></p>	<p>NHRA initiates this FSN due to the sterilization process of the mentioned lots can not be confirmed and doesn't comply with the standards.</p>
<p><b>Action should be taken</b></p>	<p>Please stop using the above mentioned lots and contact the authorized representative <b>Optica</b> at <a href="mailto:elena.feleciano@optica.net">elena.feleciano@optica.net</a> to take the necessary action for recall.</p>

**Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.**

**For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)**