


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	HotPack Vinyl / Latex Powdered Gloves
Supplier	HOT PACK TRADING CO. W.L.L
Reference	NHRA
Device picture	
Reason of Recall	Based on Circular No.(35) 2021 that prohibits the importation and distribution of powdered gloves (Vinyl/Latex) that is used in medical and non-medical fields, NHRA initiates this FSN due to the risks associated with use of powdered gloves that may cause acute respiratory tract infections, allergies and skin rashes...etc
Action should be taken	Please stop using the above mentioned gloves and contact the supplier HotPack manaf@hotpackglobal.com to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh