


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Chem tech hand sanitizer
Lot No.	All
Manufacturer	Chemistry Technical Chem Tech detergent company
Country of Origin	Bahrain
Reference	NHRA
Device picture	
Reason of Recall	NHRA initiates this FSN due to the use of Triethanolamine in the ingredients of the above-mentioned hand sanitizer which is banned by FDA due to the serious side effects when absorbed by the skin such as swelling, itching, redness, allergic reactions and skin ulcerations.
Action should be taken	Please stop using the above-mentioned hand sanitizing gel and contact the manufacturer to take the necessary action for return.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh