


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

| Device Details | |
|-------------------------|---|
| Device Name | Gripper Non-Coring Safety Needle |
| Lot No. | 3736563, 3736553, 3693997, 3675825, 3725007, 3725009, 3706171, 3725009, 3717695, 3712497, 3696519, 3694001, 3693996, 3690386, 3659762, 3674556, 3675798, 3675752, 3674138, 3671581, 3690396, 3675760, 3674130, 3653163, 3675628. |
| Manufacturer | Smiths Medical |
| Device Picture |  |
| Reference | https://www.gov.uk/drug-device-alerts/deltec-gripper-non-coring-needles-and-port-a-cath-trays-containing-gripper-needles-recall-due-to-risk-of-needle-occlusion#device-details |
| Reason of Recall | NHRA instates this FSN due to manufacturing process failure, needles may be occluded, potentially causing a delay to treatment. |
| Action taken | In case of having the above mentioned defected medical device, please contact the local agent in order to take the necessary action for replacement. |

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh