



Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details				
Device Name	Hand Sanitizing Gel			
Product Name	Lot No.	Manufacturer Name	Reason of Recall	
PIO PRO hand Sanitizer Gel	All Lots	رواد الانتاج لتصنيع الننظفات	The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. The amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer	
Antiseptic hand Sanitizing Gel		شركة مؤته الطبيه	The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. Presence of methanol.	
Diseno hand Sanitizer and moisturizer spray		مصنع المصمم العالمي للعطور	The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. Presence of methanol.	

For more information please contact Medical_Devices@nhra.bh

Action taken	Please stop using the above-mentioned defected products and contact the manufacturer to take the necessary action for recall			
Reference	https://www.sfda.gov.sa/ar/drug/news/Pages/d1-4-2020aa1.aspx			
Country of Origin	KSA			
Smscogel	All lots	الشركه السعوديه للمحاليل الطبيه	The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. The amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer	
Deter Hygienic hand gel Sanitizer		مصنع وليد علي المربعي	The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. The amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer. Presence of methanol.	

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh