



Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details					
Device Name Product Name	Hand Sanitizing Gel				
	Lot No.	Manufacturer Name	Reason of Recall		
Areej Hand Gel AntiBacterial 60ml		Unknown manufacturer	 The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. 		
Hibiclens Hand Sanitizer Gel 50ml	All Lots	Unknown manufacturer	 The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers 		
جواهر معقم لليدين Lucky Fair		مصنع مذهلة	 The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. The amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer Presence of methanol. 		

For more information please contact Medical_Devices@nhra.bh

جنفكس جل معقم لليدين	All Lots	مصنع شركة الجنيفي	 The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. Presence of methanol 	
Clean Touch Hand Sanitizer Gel		مصنع كشف الكواشف للصناعة	The amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer	
Country of Origin	KSA			
Reference	https://www.sfda.gov.sa/ar/drug/circulations/Pages/recall.aspx			
Action taken	Please stop using the above-mentioned defected products and contact the manufacturer to take the necessary action for recall			

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.