

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Olympic Brainz Monitor
Device Model	OBM00002
Serial No.	OBM00002G2119, OBM00002G2130 till OBM00002G2168
Manufacturer	Natus Medical Incorporated
Country of Origin	USA
Reason of Recall	NHRA initiates this FSN due to System functionality has not been tested on the products mentioned above, This is before exporting to distribution partners, and therefore it is not possible to confirm that the devices perform well.
Action should be taken	Please stop using the above-mentioned defected medical device and contact your authorized representative to take the necessary action to return the device for testing.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh