


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Defibrillator Electrodes
Device Serial No.	0-21-0020
Lot No.	180322-0979
Manufacturer	Schiller
Country of Origin	Switzerland
Reference	https://mhra.filecamp.com/public/file/3pnk-pe49li44
Device picture	
Reason of Recall	NHRA initiates this FSN due to potential risk that no signals or energy will be transmitted, this can result in failure to treat a patient in a life-threatening condition requiring a defibrillator shock on time.

For more information please contact Medical_Devices@nhra.bh

Action should be taken	Stop using of the affected electrodes and replace them by contacting the local agent.
-------------------------------	---

Note: In case of having the defected device model in your stock, kindly contact **Bahrain Pharmacy** at bahpharm@batelco.com.bh to take the necessary action.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh