


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	First Hand Sanitizing Gel
Lot/Batch No.	All
Manufacturer	Rayt Albayan for Disinfected and Cosmetic
Country of Origin	KSA
Reference	https://www.sfda.gov.sa/en/drug/news/Pages/d19-3-2020aa1.aspx
Device picture	
Reason of Recall	NHRA initiates this FSN due to high percentage of methyl alcohol (methanol) in the above-mentioned hand sanitizing gel which can cause bad side effects on the skin. The product is not registered in SFDA.
Action should be taken	Please stop using the above-mentioned hand sanitizing gel and dispose of it. In case of any adverse events please report to NHRA at medical_devices@nhra.bh .

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh