


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the following medical device:

Device Details	
Device Name	Nellix Endovascular Aneurysm Sealing EVAS System
Manufacturer	Endologix
Country of Origin	USA
Device picture	
Reason of Recall	This device is implanted in the body to treat the expansion of blood vessels under the kidneys (Infrarenal abdominal aortic aneurysms) and the recall is due to the failure of the treatment after two years of transplantation.
Action taken	Please contact the agent to return the product and in case of adverse event caused by the defected medical device please contact NHRA.

Note: In case of having the defected device model in your stock, kindly contact

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh