



Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

	Device Details
Device Name	Pulse Oximeter
Device Model	LK87
Lot No.	Not Available
Supplier in Bahrain	Al Sahelama General Trading
Distributed by	Jaffar Pharmacy
Manufacturer	Unknown
Reference	NHRA
Device picture	FINGERTIP PULSE OXIMETER Fingerip Pulse Oximeter Prover DO3VAA4Q5Uslin Soc. 70-100% Bpr: 25-250kpm RoHS CE TA
Reason of Recall	NHRA initiates this FSN due to the mentioned pulse oximeter was imported to Bahrain market without NHRA approval, the quality and the efficacy of the medical device is not guaranteed and the results are not accurate.
Action should be taken	Please stop using the above mentioned medical device and contact the distributor Jaffar Pharmacy at kumar@jaffar-h-ruyan.com to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh

Recall 2022 0021