

## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	VITROS CHEMISTRY PRODUCTS PERFORMANCE VERIFIER I&II
Device Model	8067324 8231474
Lot No.	Q1174 R1176
Manufacturer	QuidelOrtho
Country of Origin	USA
Reference	<a href="#">link</a>
Reason of Recall	NHRA initiates this FSN due to a Potential of mis-matched diluent lots assembled into VITROS Performance Verifier, Lots Q1174 and R1176.
Action should be taken	Please stop using the above mentioned medical device and contact the authorized representative Gulf corporation for Technology at <a href="mailto:quality@gctbahrain.com">quality@gctbahrain.com</a> to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)