

## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Voyant EA020 Electrosurgical Generator
Device Model	150779
Lot No.	200014022 200014023
Manufacturer	Applied Medical
Country of Origin	USA
Reference	<a href="#">Link</a>
Reason of Recall	NHRA initiates this FSN due to the use of an electrical board component from a non-approved manufacturer, making the units non-compliant with UL safety certifications. There is no risk to patient or user safety.
Action should be taken	Please stop using the above mentioned medical device and contact the authorized representative Wael Pharmacy Co. W.L.L at <a href="mailto:rithyl@waelpharmacy.com">rithyl@waelpharmacy.com</a> to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)