

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	S Isopropyl Alcohol
Lot No.	All
Manufacturer	S beauty group
Country of Origin	Kingdom of Bahrain
Reference	NHRA
Device picture	
Reason of Recall	NHRA initiates this FSN due to the presence of methanol which is prohibited to be used as an ingredient in hand sanitizer due to its toxic effects and the amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer and as a result will impact its effectiveness against COVID-19.
Action should be taken	Please stop using the above-mentioned product and contact the supplier to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh