

## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	SEPTIRAZ isopropyl alcohol 70% 250 ML Spray
Lot No.	20F52L
Manufacturer	مصنع رازي المدينة للصناعات الدوائية Al Razi Pharmaceutical Company
Country of Origin	KSA
Reference	<a href="https://www.sfda.gov.sa/sites/default/files/2020-07/2109_001.pdf">https://www.sfda.gov.sa/sites/default/files/2020-07/2109_001.pdf</a>
Device picture	
Reason of Recall	NHRA initiates this FSN due to the amount of sterile alcohol (Isopropanol) is less than 70% which is considered the minimum in hand sanitizer.
Action should be taken	Please stop using the above-mentioned defected products and contact the manufacturer to take the necessary action for recall

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)