


## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
<b>Device Name</b>	Medical Latex Examination Gloves
<b>Device Model</b>	Transo-glove Medical Latex Examination Gloves
<b>Lot No.</b>	ALL Lots
<b>Manufacturer</b>	Unknown
<b>Device picture</b>	
<b>Reason of Recall</b>	Gloves contains powder which is not stated on the product label and the use of powder gloves is prohibited in the kingdom of Bahrain due to potential risks on users, moreover as the manufacturer name is not printed on the label this gloves is considered to be from an unknown source and its quality and safety cannot be guaranteed.
<b>Action should be taken</b>	Kindly contact the local importer to take the necessary actions to recall/withdraw the above-mentioned Medical Device, if you faced any side effects after using the product please contact us on the email below
<b>Local Authorized Representative</b>	Behzad Medical Establishment W.L.L.

**Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.**

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)