


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	Valera 200 MI
Lot No.	20031272
Manufacturer	UNI care
Reference	https://www.sfda.gov.sa/sites/default/files/2020-08/2159_001.pdf
Device picture	
Reason of Recall	NHRA initiates this FSN due to the amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers.
Action should be taken	Please stop using the above-mentioned defected products and contact the manufacturer to take the necessary action for recall

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh