


Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

| Device Details | |
|-------------------------------|--|
| Device Name | Vital Oxide Disinfectant |
| Lot No. | All |
| Manufacturer | Vin-Technology Systems W.L.L (Vinner) |
| Country of Origin | Bahrain |
| Reference | NHRA |
| Device picture |  |
| Reason of Recall | NHRA initiates this FSN because the above-mentioned disinfectant is being re-filled and re-packaged in a facility that is “not licensed” which may affect the safety, quality and efficacy of the product. |
| Action should be taken | Please stop using the above disinfectant and contact the seller to take the necessary action for recall. |

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh