

## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	A Clean Alcohol
Lot No.	042
Manufacturer	شركة مصنع فذ السعودية للمنظفات والمطهرات
Country of Origin	KSA
Reference	SFDA
Reason of Recall	The amount of sterile alcohol (Ethanol) is less than 60%, which is the minimum in hand sanitizers. Presence of methanol.
Action should be taken	Stop using the above-mentioned had sanitizer and contact the supplier to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)