

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	MRI COMPACTIBLE VENTILATOR
Device Model	ZOLL Z VENT 731 SERIES
Serial Number	(21)AY22K049977
Manufacturer	Zoll Medical Corporation
Country of Origin	USA
Reference	Link
Reason of Alert	NHRA initiates this FSN due to identifying that the Operator's Guides and Quick Reference Guides for MRI Compatible 731 ZOLL Ventilators incorrectly stated the ventilator placement.
Action should be taken	Please refer to "Actions to be taken by Customer/ User" in the attached FSN And for more information please contact the authorized representative Gulf corporation for Technology W.L.L at quality@gctbahrain.com

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh