

Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details						
Device Name	BD BBL™ Sensi-Disc™					
Catalogue Number	231251	231539	231633	231660	231682	232175
	231263	231544	231635	231664	231692	232219
	231264	231607	231641	231672	231696	
	231274	231621	231645	231674	231704	
	231344	231629	231658	231678	231706	
Manufacturer	Becton, Dickinson and Company (BD)					
Country of Origin	USA					
Reference	Link					
Reason of Alert	NHRA initiates this FSN due to identifying that BD BBL™ Sensi-Disc™ may produce variable results for Haemophilus susceptibility testing, leading to potential inaccuracies.					
Action should be taken	Please refer to “Actions to be taken by Customer/ User” in the attached FSN And for more information please contact the authorized representative Wael Pharmacy Co. W.L.L at sharon@waelpharmacy.com					

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact Medical_Devices@nhra.bh