

## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	CENTRICITY RIS -i
Model No.	Software Version 7.0.4
Manufacturer	GE HEALTHCARE
Country of Origin	USA
Reference	<a href="#">Link</a>
Reason of Alert	A cybersecurity vulnerability affecting Centricity RIS-i configured with Instant Messenger.
Action should be taken	Please refer to “Actions to be taken by Customer/ User” in the attached FSN And for more information please contact the authorized representative GULF CORPORATION FOR TECHNOLOGY W.L.L at <a href="mailto:Quality@gctbahrain.com">Quality@gctbahrain.com</a> .

**Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.**

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)