

## Medical Devices Safety Notice

The National Health Regulatory Authority would like to alert all governmental and private healthcare facilities, local agents and distributors that the below medical device:

Device Details	
Device Name	BD BACTEC™ MGIT™ 960 PZA Kit
Device Model	245128
Lot No.	4262044 4362418
Manufacturer	Becton, Dickinson & Company
Country of Origin	USA
Reference	<a href="#">Link</a>
Reason of Recall	NHRA initiates this FSN due to the risk of false pyrazinamide (PZA) resistance results, which may lead to suboptimal tuberculosis treatment, prolonged therapy, and increased risk of side effects such as hepatotoxicity and neuropathy.
Action should be taken	Please stop using the above mentioned medical device and contact the authorized representative Wael Pharmacy Co. at <a href="mailto:sharon@waelpharmacy.com">sharon@waelpharmacy.com</a> to take the necessary action for recall.

Your cooperation is highly appreciated in improving health services in the Kingdom of Bahrain.

For more information please contact [Medical\\_Devices@nhra.bh](mailto:Medical_Devices@nhra.bh)